




SHIELDskin Xtreme Sterile Orange Nitrile 300 DI

pure¹¹-Nr.: 05209, Hersteller: Shield Scientific



Zusammenfassung

- Neue pure11-Artikelnummer (ab 01.07.2023): 1105209
- Material: Nitril/Neopren
- Handspezifisch
- Puderfrei
- Latexfrei
- AQL-Wert (Acceptable Quality Level): 0.65
- Gammasterilisiert
- Texturierte Handinnenflächen und Fingerspitzen
- Vulkanisationsbeschleunigerfrei
- Eignet sich perfekt als Unterziehhandschuh
- Einzigartige Skin Nitrile™ Technologie
- Mit weißer Innenseite
- Einfache Wandstärke 0,15 mm (Mittelfinger)
- Reduziertes Allergierisiko (Type IV)
- Viren- und mikroorganismenresistent
- Gut geeignet zum Double-Gloving
- Beständig gegen eine Vielzahl von Zytostatika
- Gute Chemikalienbeständigkeit

Empfohlene Reinraumklassen

| | | | | | | | | |
|-----|--------------------------|--------------------------|-------------------------------------|-------------------------------------|-------------------------------------|-------------------------------------|-------------------------------------|-------------------------------------|
| ISO | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> |
| GMP | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

Produktvarianten

pure¹¹-Nr.: 052095b

Farbe: Orange / Größe: 5,5 / Herst.-Nr.: 696551 / VE: 200 Paar

pure¹¹-Nr.: 052096

Farbe: Orange / Größe: 6,0 / Herst.-Nr.: 696552 / VE: 200 Paar

pure¹¹-Nr.: 052096b

Farbe: Orange / Größe: 6,5 / Herst.-Nr.: 696553 / VE: 200 Paar

pure¹¹-Nr.: 052097

Farbe: Orange / Größe: 7,0 / Herst.-Nr.: 696554 / VE: 200 Paar

pure¹¹-Nr.: 052097b

Farbe: Orange / Größe: 7,5 / Herst.-Nr.: 696555 / VE: 200 Paar

pure¹¹-Nr.: 052098

Farbe: Orange / Größe: 8,0 / Herst.-Nr.: 696556 / VE: 200 Paar

pure¹¹-Nr.: 052098b

Farbe: Orange / Größe: 8,5 / Herst.-Nr.: 696557 / VE: 200 Paar

pure¹¹-Nr.: 052099

Farbe: Orange / Größe: 9,0 / Herst.-Nr.: 696558 / VE: 200 Paar

pure¹¹-Nr.: 0520910

Farbe: Orange / Größe: 10,0 / Herst.-Nr.: 696559 / VE: 200 Paar

Quelle: <https://www.pure11.de/shieldskin-xtreme-sterile-orange-nitrile-300-di>



SHIELDskin XTREME™
A REVOLUTION IN GLOVE TECHNOLOGY

Sterile

BIO
CONTAMINATION CONTROL

SHIELDskin XTREME™

Sterile ORANGE NITRILE™ 300 DI





Sterile

Bio contamination control

DI

Basic contamination control

- ⇒ Powder-free single DI washed hand-specific standard length (300 mm / 11.8") sterile nitrile/neoprene cleanroom gloves.
- ⇒ Personal Protective Equipment Category III (PPE - Complex Design) according to Regulation (EU) 2016/425.
- ⇒ Fully compliant to the latest PPE Protective gloves EU norms against chemicals, micro-organisms and viruses.

| DESCRIPTION | |
|-------------|---|
| Formulation | Nitrile and neoprene synthetic rubber (<i>acrylonitrile butadiene and polychloroprene</i>). |
| Design | Orange, hand-specific, beaded cuff, textured palm and fingers. |
| Packaging | 1 pair per PE peel pouch - 20 pouches per sealed poly bag - 10 poly bags per PE bag per carton. |

| SIZES | 5.5 | 6.0 | 6.5 | 7.0 | 7.5 | 8.0 | 8.5 | 9 | 10 |
|-------|---------|---------|---------|---------|---------|---------|---------|---------|---------|
| Codes | 69 6551 | 69 6552 | 69 6553 | 69 6554 | 69 6555 | 69 6556 | 69 6557 | 69 6558 | 69 6559 |

| STANDARDS | |
|---------------------------|---|
| CE registration | PPE Category III (Complex Design) - Regulation (EU) 2016/425. Notified Body No 0598: SGS Fimko Oy, Helsinki - FINLAND. |
| EU PPE norms | ISO 21420:2020, ISO 374-1:2016+A1:2018, ISO 374-2:2019, ISO 374-4:2019, ISO 374-5:2016, EN 16523-1:2015+A1:2018 and ISO 16604:2004 Procedure B. |
| EU MDR norms ¹ | EN 455-1:2000, EN 455-2:2015, EN 455-3:2015 and EN 455-4:2009. |
| USA standards | ASTM D3767-03 (2020), ASTM D573-04 (2019), ASTM D412-16, ASTM D6978-05 (2019) and IEST-RP-CC005.4 (2013). |
| Other standards | ISO 11137-2:2015, ISO 10993-10:2010. |

¹With reference to Regulation (EU) 2017/425 for Medical Devices

| QUALITY | |
|-------------------|---|
| Quality assurance | Production management in accordance with ISO 9001:2015 and ISO 13485:2016. |
| Technology | uniSHIELD™ single-walled protection to offer an ideal compromise between comfort and protection. Synthetic soft polymer, based on Skin Nitrile™ technology with a blend of polychloroprene. Compatible with sterile processing environments due to paperless packaging and multiple post leaching of gloves (single washed in deionised water). |

| DOCUMENTATION | |
|---------------------------------|--|
| Declaration of conformity | These documents can be freely downloaded from the product page on our website: www.shieldscientific.com . |
| EU type examination certificate | For easy access, scan the QR code. |
| User's instructions | |
| Certificate of conformance | To access CoC and CoI, you need to be registered. Please contact us at info@shieldscientific.com or call your SHIELD Scientific representative. |
| Certificate of irradiation | |



PHYSICAL PROPERTIES



| NOMINAL THICKNESS | mm ² | mil | Norm |
|-------------------|-----------------|-----|----------------------|
| ⇒ Finger | 0.15 | 5.9 | ASTM D3767-03 (2020) |
| ⇒ Palm | 0.14 | 5.5 | |
| ⇒ Cuff | 0.09 | 3.5 | |

² Thickness (+/- 0.03 mm)

| LENGTH | Minimum | Typical | Norm |
|--|------------------|----------------|----------------|
| ⇒ From middle finger tip to edge of cuff | ≥ 300 mm / 11.8" | 305 mm / 12.0" | ISO 21420:2020 |

| STRENGTH PROPERTIES | Force at break (spec.) | | Ultimate elongation (spec.) | Force at break (typical) | Norm |
|---------------------|------------------------|--------|-----------------------------|--------------------------|--|
| ⇒ Before aging | ≥ 6.0N | 14 MPa | ≥ 500% | 12.0N | EN 455-2:2015 ASTM D573-04 (2019) & ASTM D412-16 |
| ⇒ After aging | ≥ 6.0N | 14 MPa | ≥ 400% | 11.0N | |

| FREEDOM FROM HOLES | Performance | Norm |
|----------------------------------|-------------------------------|---------------------------------|
| ⇒ Acceptable Quality Level (AQL) | < 0.65 ³ - Level 3 | ISO 374-2:2019 EN 455-1:2000 |

³ AQL as defined per ISO 2859-1:1999 for sampling by attributes.

| RISKS | Description | Norm |
|-----------------|--|---|
| Micro-organisms | 1000 ml water test. Performance level 3, AQL < 0.65 (inspection level G1). | ISO 374-2:2019 |
| Viruses | Viral penetration test using Phi-X174 bacteriophage according to ISO 16604:2004 Procedure B. | ISO 374-5:2016 |
| Chemicals | <u>Performance</u> : Type B (KPT). <u>Permeation</u> : Extensively tested. Online chemical resistance guide on www.shieldscientific.com . <u>Degradation</u> : Tested for determination of resistance to degradation by chemicals. | ISO 374-1:2016+A1:2018 EN 16523-1:2015+A1:2018 ISO 374-4:2019 |
| Cytotoxic | Tested for permeation to potentially hazardous cancer chemotherapy drugs under conditions of continuous contact. | ASTM D6978-05 (2019) |

CLEANLINESS PROPERTIES

| PARTICLES | Specification | Typical value | Test method |
|-----------------------------------|---------------------------|-----------------|-----------------|
| Particles/cm ² ≥ 0.5µm | < 3,000 particles (spec.) | 1,000 particles | IEST-RP-CC005.4 |

| EXTRACTABLES (ION) | Specification (µg/cm ²) | Typical value (µg/cm ²) | Test method |
|------------------------------|-------------------------------------|-------------------------------------|-----------------|
| Ammonium (NH ₄) | 0.050 | 0.015 | IEST-RP-CC005.4 |
| Bromide (Br) | 0.030 | < 0.008 | |
| Calcium (Ca) | 0.500 | 0.300 | |
| Chloride (Cl) | 0.400 | 0.100 | |
| Fluoride (F) | 0.010 | < 0.008 | |
| Magnesium (Mg) | 0.010 | < 0.008 | |
| Nitrate (NO ₃) | 0.200 | 0.090 | |
| Nitrite (NO ₂) | 0.050 | < 0.008 | |
| Phosphate (PO ₄) | 0.050 | < 0.008 | |
| Potassium (K) | 0.050 | 0.020 | |
| Sodium (Na) | 0.050 | 0.008 | |
| Sulphate (SO ₄) | 0.050 | 0.008 | |

| EXTRA TESTS | Description | Test method |
|-------------|---|-------------------|
| Sterility | Terminally sterilized by gamma irradiation to Sterility Assurance Level (SAL) of 10 ⁻⁶ (ISO 11137-2:2015). | |
| Endotoxins | Low Endotoxin content at < 20 EU/pair demonstrated by Limulus Amoebocyte Lysate (LAL) kinetic turbidimetric test. | EN 455-3:2015 |
| NVR | Maximum 30 µg/g. | IEST-RP-CC005.4 |
| FTIR | Non-detectable levels of silicone, amide and DOP. | IEST-RP-CC005.4 |
| ESD | Tested for electrostatic properties. | EN 1149-1/2/3 & 5 |

| ALLERGIES | |
|--------------------|---|
| Bio-Compatibility | Demonstrated by skin irritation and sensitization tests in accordance with ISO 10993-10:2010. |
| Accelerators | Accelerator-free to minimize the risk of allergic contact dermatitis (also known as Type IV, delayed hypersensitivity or chemical allergy). |
| Chemical Allergens | Non-detectable levels using aqueous solution extraction (Phosphate buffered solution) and High Performance Liquid Chromatography (HPLC) assay method for quantitative analysis. |
| Latex Protein | Latex-free. |

CHEMICAL RESISTANCE GUIDE



SHIELDskin XTREME* Sterile ORANGE NITRILE* 300 DI



- Category III PPE glove (PPE Regulation (EU) 2016/425)
- Complex Design - For mortal & irreversible risks
- Powder-free orange nitrile/polychloroprene glove
- Hand-specific
- 300 mm / 0.14 mm (EN 420:2003+A1:2009)
- Biological risk (ISO 374-5:2016 VIRUS)
- AQL 0.65 (EN 374-2:2014 Level 3)
- Viral penetration test (ISO 16604:2004 Procedure B)
- Chemical risk (ISO 374-1:2016+A1:2018 - Type B KPT)
- Waterproof and for low chemical protection
- Tested for chemical permeation (EN 16523-1:2015+A1:2018)
- Typical particle levels: less than 1000 per cm² more or equal 0.5µm
- Type I hypersensitivity eliminated - Type IV hypersensitivity reduced

| | |
|-------------------------------------|--------------------|
| 64-19-7 Acetic Acid 100% | LEVEL 0 6 min |
| 10127-02-3 Acridine orange | LEVEL 6 480 min |
| 79-06-1 Acrylamide 40% | LEVEL 6 480 min |
| 79-10-7 Acrylic acid 99% | LEVEL 0 4 min |
| 107-13-1 Acrylonitrile 99% | LEVEL 0 0 min |
| 1336-21-6 Ammonium Hydroxide 25% | LEVEL 1 28 min |
| 62-53-3 Aniline 99,9% | LEVEL 1 14 min |

| | |
|--|--------------------|
| 100-51-6 Benzyl Alcohol | LEVEL 0 9 min |
| 74-97-5 Bromochloromethane | LEVEL 3 66 min |
| 71-36-3 Butanol 100% | LEVEL 2 39 min |
| 97-88-1 Butyl methacrylate 99,9% | LEVEL 1 11 min |
| 67-66-3 Chloroform 99,8% | LEVEL 0 0 min |
| 77-92-9 Citric acid 30% | LEVEL 6 480 min |
| 110-82-7 Cyclohexane | LEVEL 6 480 min |
| 108-94-1 Cyclohexanone 99% | LEVEL 0 6 min |
| 75-09-2 Dichloromethane 99% | LEVEL 0 0 min |
| 127-19-5 Dimethyl Acetamide 99% | LEVEL 0 8 min |
| 68-12-2 Dimethyl Formamide 99% | LEVEL 0 4 min |
| 67-68-5 Dimethyl Sulfoxide 99% (DMSO) | LEVEL 2 42 min |
| 64-17-5 Ethanol 99.8% | LEVEL 1 16 min |
| 64-17-5 Ethanol 70% | LEVEL 2 31 min |

| | |
|--|--------------------|
| 1239-45-8 Ethidium Bromide 5% | LEVEL 6 480 min |
| 107-21-1 Ethylene Glycol | LEVEL 0 0 min |
| 50-00-0 Formaldehyde 10% | LEVEL 6 480 min |
| 50-00-0 Formaldehyde 37% | LEVEL 6 480 min |
| 75-12-7 Formamide 99% | LEVEL 3 99 min |
| 64-18-6 Formic acid 98,5% | LEVEL 0 3 min |
| 111-30-8 Glutaraldehyde 25% | LEVEL 6 480 min |
| 111-30-8 Glutaraldehyde 2.5% | LEVEL 6 480 min |
| 50-01-1 Guanidine Hydrochloride | LEVEL 6 480 min |
| 7803-57-8 Hydrazine monohydrate 80% | LEVEL 6 480 min |
| 7803-57-8 Hydrazine monohydrate 98% | LEVEL 4 180 min |
| 7647-01-0 Hydrochloric Acid 37% | LEVEL 3 116 min |
| 7664-39-3 Hydrofluoric Acid 40% | LEVEL 1 10 min |
| 7664-39-3 Hydrofluoric Acid 48% | LEVEL 0 6 min |


| | |
|--|------------------------------|
| 7664-39-3 Hydrogen Fluoride 48% | LEVEL 0 6 min |
| 7722-84-1 Hydrogen Peroxide 30% | LEVEL 6 480 min DR 22% |
| 7722-84-1 Hydrogen Peroxide 12% | LEVEL 6 480 min |
| 78-83-1 Isobutanol 99% | LEVEL 3 76 min |
| 540-84-1 Iso-Octane 99% | LEVEL 6 480 min |
| 67-63-0 Isopropanol 100% | LEVEL 2 43 min |
| 67-63-0 Isopropanol 70% | LEVEL 3 61 min |
| Mixed Solution Klercide Premier - WFI 60/40 sterile Alcohol | LEVEL 2 41 min |
| Mixed Solution Klercide Premier - WFI 70/30 sterile IPA | LEVEL 3 106 min |
| Mixed Solution Klercide 70/30 sterile IPA | LEVEL 3 76 min |
| 108-39-4 m-Cresol 98.5% | LEVEL 2 59 min |
| 67-56-1 Methanol 99,9% | LEVEL 0 5 min |
| 37143-54-7 1-Methoxy-2-propylamine 95% | LEVEL 0 3 min |
| 108-87-2 Methylcyclohexane 99,9% | LEVEL 2 55 min |

| | |
|--|--------------------|
| 1634-04-4 Methyl Tert Butyl Esther (MTBE) | LEVEL 1 11 min |
| 96-47-9 2-Methyltetrahydrofuran 99,9% | LEVEL 0 1 min |
| Mixed Solution Mucocit®-T branded mixture | LEVEL 6 480 min |
| 110-54-3 n-Hexane 95% | LEVEL 3 97 min |
| 54-11-5 Nicotine 98% | LEVEL 4 151 min |
| 7697-37-2 Nitric Acid, 50% | LEVEL 2 49 min |
| 109-66-0 n-Pentane 98% | LEVEL 2 56 min |
| Mixed Solution Perform sterile concentrate Oxy 2% | LEVEL 6 480 min |
| Mixed Solution Perform sterile concentrate PAA 3% | LEVEL 6 480 min |
| Mixed Solution Perform sterile PAA ready to use | LEVEL 6 480 min |
| 108-95-2 Phenol 50% | LEVEL 1 24 min |
| 7664-38-2 Phosphoric Acid, 30% | LEVEL 6 480 min |
| 7664-38-2 Phosphoric acid, 85% | LEVEL 6 480 min |
| 1310-58-3 Potassium Hydroxide 40% | LEVEL 6 480 min |

| | |
|---|--------------------|
| 75-56-9 Propylene oxide 99% | LEVEL 0 0 min |
| 110-86-1 Pyridine | LEVEL 0 1 min |
| 598-75-4 Secondary isoamyl alcohol 98% | LEVEL 2 55 min |
| 127-09-3 Sodium acetate Sat. solution | LEVEL 6 480 min |
| 1310-73-2 Sodium Hydroxide 40% | LEVEL 6 480 min |
| 1310-73-2 Sodium Hydroxide, 50% | LEVEL 6 480 min |
| 7681-52-9 Sodium Hypochlorite 13% | LEVEL 6 480 min |
| Mixed Solution SPOR-KLENZ Solution | LEVEL 6 480 min |
| 100-42-5 Styrene 99.9% | LEVEL 0 0 min |
| 7664-93-9 Sulphuric Acid 10% | LEVEL 6 480 min |
| 7664-93-9 Sulphuric Acid 95%-98% | LEVEL 0 6 min |
| 7664-93-9 Sulphuric Acid 50% | LEVEL 6 480 min |
| 109-99-9 Tetrahydrofuran 99.9% | LEVEL 0 0 min |
| 76-03-9 Trichloroacetic acid 10% | LEVEL 6 480 min |

| | |
|--|---------------------------|
| 121-44-8 Triethylamine 99% | LEVEL 2 36 min |
| 95-63-6 1,2,4- Trimethylbenzene 98% | LEVEL 1 13 min |
| 108-67-8 1,3,5-Trimethylbenzene 98% | LEVEL 1 10 min |
| 77-86-1 Tris (hydroxymethyl) aminomethane Sat. solution | LEVEL 6 480 min |
| 1330-20-7 Xylene 98,5% | LEVEL 0 2 min |

DISCLAIMER: The data provided was based on gloves tested under laboratory conditions, in accordance with EN 16523-1:2015 (formerly EN 374-3:2003) and EN 374-4:2013. The information is for guidance only and may not reflect the user's application. A risk assessment should always be made by purchaser to assess the suitability of gloves for a specific application.





EU DECLARATION OF CONFORMITY

FOR PERSONAL PROTECTIVE EQUIPEMENT

Originator: J.F ROBLES

Revision: 6

Revision date: 31.12.2019

Validity date: 31.12.2024

| | |
|-----------------------|---|
| PRODUCT | SHIELDskin XTREME™ Sterile ORANGE NITRILE™ 300 DI |
| DESCRIPTION | Powder Free DI washed Pair-packed Hand-specific 30cm Cleanroom Gloves |
| CLASSIFICATION | Personal Protective Equipment (PPE) Category III (Complex Design) |

| SHIELD Scientific codes | Sizes |
|-------------------------|-------|
| 69 6551 | 5.5 |
| 69 6552 | 6 |
| 69 6553 | 6.5 |
| 69 6554 | 7 |
| 69 6555 | 7.5 |
| 69 6556 | 8 |
| 69 6557 | 8.5 |
| 69 6558 | 9 |
| 69 6559 | 10 |

The manufacturer established in the Union:

SHIELD Scientific B.V.

Dr Willem Dreeslaan 1 – 6721 ND BENNEKOM – THE NETHERLANDS

declares under his/her sole responsibility that the PPE (product codes as mentioned above) described hereafter:

SHIELDskin XTREME™ Sterile ORANGE NITRILE™ 300 DI

is in conformity with the provisions of Regulation (EU) 2016/425 and with the harmonized standards EN ISO 374-1:2016 (as a Type B glove against reagents: K, P & T), EN 374-2:2014 (performance level 3, including protection against viruses), EN 16523-1:2015, EN 374-4:2013, EN ISO 374-5:2016 and EN 420:2003 + A1:2009. This device is identical to the PPE, which is the subject of EU Type Examination (Module B) certificate of conformity *no. FI12/123456 (Technical file submitted/Notified Body answer pending)* issued by the Notified Body:

SGS FIMKO OY (Notified Body No: 0598)

Särkiniementie 3 - 00211 Helsinki - Finland

This device is subject to the procedure set out in Article VIII (Module D) of the Regulation under the surveillance of the Notified Body:

SGS FIMKO OY (Notified Body No: 0598)

Särkiniementie 3 - 00211 Helsinki - Finland

Signed for and behalf of SHIELD Scientific B.V



J.F ROBLES
General Manager

Date: 31th December 2019

Place: Bennekom

Validity of this declaration: 31th December 2019 until 31th December 2024



CHEMICAL ANALYTICAL SERVICES

TEST REPORT ASTM D 6978-05

Project No: 104597B

SUBJECT: Permeation testing per ASTM D 6978-05 on sample submitted

SAMPLE: SHIELDskin XTREME™ Sterile ORANGE NITRILE™ 300 DI

Test Laboratory: Akron Rubber Development Laboratories Inc
2887 Gilchrist Road
Akron Ohio 44305
UNITED STATES

TEST CHEMICALS:

Test chemicals and their sources are listed in Table 1.

Table 1. List of the Test Chemicals, Sources, and Expiration Dates

| Test Chemical | Chemical Source |
|----------------------------|-----------------|
| Carmustine (BCNU) | Bristol-Myers |
| Cyclophosphamide (Cytosan) | Sigma |
| Doxorubicin Hydrochloride | Pfizer |
| Etoposide (Toposar) | APP |
| Fluorouracil | APP |
| Paclitaxel (Taxol) | Teva |
| Thiotepa | Sigma Aldrich |

COLLECTION MEDIA:

The collection media which were selected are listed in Table 2.

Table 2. Collection Media for Test Chemicals

| Test Chemical | Concentration | Collection Medium |
|----------------------------|---------------|-----------------------------------|
| Carmustine (BCNU) | 3,300 ppm | 10% Ethanol Aqueous Solution |
| Cyclophosphamide (Cytocan) | 20,000 ppm | Distilled Water |
| Doxorubicin Hydrochloride | 2,000 ppm | Distilled Water |
| Etoposide (Toposar) | 20,000 ppm | Distilled Water |
| Fluorouracil | 50,000 ppm | 9.20 pH Sodium Hydroxide Solution |
| Paclitaxel (Taxol) | 6,000 ppm | 30% Methanol Aqueous Solution |
| Thiotepa | 10,000 ppm | Distilled Water |

NOTE: The chemotherapy drugs were prepared as required by the ASTM D 6978-05 Standard at 5.2.2.5, page 2, using the highest concentration of the drugs to which a healthcare worker might be exposed during handling as referenced in the most recent edition of the Physician's Desk Reference or the package inserts of the testing drugs.

TESTING CONDITIONS:

| | |
|--------------------------------------|--|
| Standard Test Method Used: | ASTM D 6978-05 |
| Deviation From Standard Test Method: | Used 1" Permeation Cell |
| Analytical Method: | UV/VIS Spectrometry |
| Testing Temperature: | 35.0°C ± 2.0 |
| Collection System: | Closed Loop |
| Specimen Area Exposed: | 5.067 cm ² |
| Selected Data Points: | 25/test |
| Number of Specimens Tested: | 3/test |
| Location Sampled From: | Cuff area |
| Comments/Other Conditions: | Magnetic stir bar was used in the sampling chamber |

DETECTION METHOD OF CHEMICAL PERMEATION; UV/VIS ABSORPTION SPECTROMETRY

Instrument: Perkin Elmer UV/VIS Spectrometer Lambda 25

UV/VIS Absorption Spectrometry was used to measure the absorbance of test chemicals which permeated through the specimens into the collection medium. The collection medium was circulated in a closed loop at 11 ml/minute of flow rate through the testing period. Data collection was performed according to the programmed schedule by means of UV Winlab software from the Perkin Elmer Corporation. The list of the characteristics wavelengths is shown in Table 3.

Table 3. Characteristics Wavelengths used in UV/VIS Absorption Spectrometry.

| Test Chemical | Wave length (nm) |
|----------------------------|------------------|
| Carmustine (BCNU) | 229 |
| Cyclophosphamide (Cytocan) | 200 |
| Doxorubicin Hydrochloride | 232 |
| Etoposide (Toposar) | 205 |
| Fluorouracil | 269 |
| Paclitaxel (Taxol) | 231 |
| Thiotepa | 199 |

SAMPLE CHARACTERISTICS

Table 4. Sample Characteristics

| SAMPLE | THICKNESS (mm) | WEIGHT/UNIT AREA (g/m ²) |
|----------------------------|----------------|--------------------------------------|
| | Average | |
| Carmustine | 0.089 | 88.6 |
| Cyclophosphamide (Cytoxan) | 0.090 | 88.6 |
| Doxorubicin Hydrochloride | 0.089 | 88.6 |
| Etoposide (Toposar) | 0.086 | 88.6 |
| Fluorouracil | 0.088 | 88.6 |
| Paclitaxel (Taxol) | 0.090 | 88.6 |
| Thiotepa | 0.091 | 88.6 |

CALCULATION:

Average permeation rate for each sampling time interval (except time = 0 minutes) was calculated using the equation below:

$$P_i = ((C_i - C_{i-1}) \times V_i) / ((T_i - T_{i-1}) \times A)$$

Where:

- P_i = average permeation rate, ug/cm²/min. at time interval T_i - T_{i-1}
- C_i = concentration of test chemical detected in collection medium, ug/l, at time T_i
- V_i = volume of collection medium, l
- i = an index number starting with i = 1 for the first sample
- T_i = sampling time, minutes
- A = area of specimen in contact with the test chemical, cm²

RESULTS

The permeation test results are listed in Table 5:

Table 5 Permeation Test Results:

| Test Chemical | Breakthrough* Detection Time | Steady State Permeation Rate |
|---|--|--|
| | Minutes | (Avg.) $\mu\text{g}/\text{cm}^2/\text{min.}$ |
| Carmustine (BCNU) 3,300 ppm / 3.3 mg/ml | 30.8 (46.6, 35.5, 30.8) | 0.7 (0.6, 0.9, 0.7) |
| Cyclophosphamide (Cytosan) 20,000 ppm / 20.0 mg/ml | No breakthrough was detected up to 240 minutes | N/A |
| Doxorubicin Hydrochloride 2,000 ppm / 2.0 mg/ml | No breakthrough was detected up to 240 minutes | N/A |
| Etoposide (Toposar) 20,000 ppm / 20.0 mg/ml | No breakthrough was detected up to 240 minutes | N/A |
| Fluorouracil 50,000 ppm / 50.0 mg/ml | No breakthrough was detected up to 240 minutes | N/A |
| Paclitaxel (Taxol) 6,000 ppm / 6.0 mg/ml | No breakthrough was detected up to 240 minutes | N/A |
| Thiotepa 10,000 ppm / 10.0 mg/ml | 106.0 (106.0, 136.6, 121.2) | 0.2 (0.2, 0.1, 0.2) |

*Breakthrough detection time is the time in minutes measured from the start of the test to the sampling time that immediately precedes the sampling time at which the permeation rate reaches $0.01 \mu\text{g}/\text{cm}^2/\text{min.}$

SHIELD Scientific B.V.



A handwritten signature in blue ink, appearing to read 'Cisco Robles', is written over the SHIELD Scientific logo.

Cisco Robles
General Manager